

Data Path : W:\HPCHEM1\MSVOA X\Data\VX050418\
 Data File : VX001391.D
 Acq On : 04 May 2018 11:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 04 14:28:10 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.553	0.540	2.4	78	0.00
3 P	Chloromethane	0.543	0.508	6.4	82	0.00
4 C	Vinyl Chloride	0.572	0.555	3.0#	81	0.00
5 T	Bromomethane	0.486	0.410	15.6	76	0.00
6 T	Chloroethane	0.361	0.374	-3.6	89	0.00
7 T	Trichlorofluoromethane	0.965	1.005	-4.1	88	0.00
8 T	Diethyl Ether	0.331	0.357	-7.9	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.556	0.570	-2.5	88	0.00
10 T	Methyl Iodide	0.559	0.829	-48.3#	118	0.00
11 T	Tert butyl alcohol	0.119	0.149	-25.2#	103	0.02
12 CM	1,1-Dichloroethene	0.502	0.501	0.2#	86	0.00
13 T	Acrolein	0.069	0.039	43.5#	48#	0.00
14 T	Allyl chloride	0.877	0.935	-6.6	89	0.00
15 T	Acrylonitrile	0.276	0.317	-14.9	98	0.00
16 T	Acetone	0.295	0.309	-4.7	89	0.00
17 T	Carbon Disulfide	1.307	1.121	14.2	75	0.00
18 T	Methyl Acetate	0.698	0.874	-25.2#	104	0.00
19 T	Methyl tert-butyl Ether	1.708	1.829	-7.1	91	0.00
20 T	Methylene Chloride	0.639	0.565	11.6	87	0.00
21 T	trans-1,2-Dichloroethene	0.546	0.529	3.1	85	0.00
22 T	Diisopropyl ether	1.690	1.631	3.5	82	0.00
23 T	Vinyl Acetate	1.432	1.485	-3.7	87	0.00
24 P	1,1-Dichloroethane	0.941	1.015	-7.9	93	0.00
25 T	2-Butanone	0.399	0.421	-5.5	89	0.01
26 T	2,2-Dichloropropane	0.766	0.730	4.7	81	0.00
27 T	cis-1,2-Dichloroethene	0.600	0.572	4.7	83	0.00
28 T	Bromochloromethane	0.423	0.467	-10.4	87	0.00
29 T	Tetrahydrofuran	0.246	0.271	-10.2	93	0.00
30 C	Chloroform	0.995	0.986	0.9#	85	0.00
31 T	Cyclohexane	0.850	0.761	10.5	78	0.00
32 T	1,1,1-Trichloroethane	0.870	0.866	0.5	83	0.00
33 S	1,2-Dichloroethane-d4	0.692	0.650	6.1	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.431	0.437	-1.4	85	0.00
36 T	1,1-Dichloropropene	0.577	0.533	7.6	80	0.00
37 T	Ethyl Acetate	0.583	0.624	-7.0	90	0.01
38 T	Carbon Tetrachloride	0.605	0.600	0.8	82	0.00
39 T	Methylcyclohexane	0.683	0.629	7.9	78	0.00
40 TM	Benzene	1.660	1.621	2.3	82	0.00
41 T	Methacrylonitrile	0.339	0.348	-2.7	89	0.00
42 TM	1,2-Dichloroethane	0.643	0.618	3.9	83	0.00
43 T	Isopropyl Acetate	0.943	0.995	-5.5	85	0.00
44 TM	Trichloroethene	0.508	0.500	1.6	84	0.00
45 C	1,2-Dichloropropane	0.426	0.419	1.6#	82	0.00

Data Path : W:\HPCHEM1\MSVOA X\Data\VX050418\
 Data File : VX001391.D
 Acq On : 04 May 2018 11:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 04 14:28:10 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 02 01:48:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.299	0.299	0.0	85	0.00
47 T	Bromodichloromethane	0.534	0.551	-3.2	83	0.00
48 T	Methyl methacrylate	0.501	0.523	-4.4	86	0.00
49 T	1,4-Dioxane	0.009	0.011	-22.2#	97	0.00
50 S	Toluene-d8	1.539	1.600	-4.0	82	0.00
51 T	4-Methyl-2-Pentanone	0.608	0.692	-13.8	91	0.00
52 CM	Toluene	1.085	1.083	0.2#	83	0.00
53 T	t-1,3-Dichloropropene	0.635	0.651	-2.5	82	0.00
54 T	cis-1,3-Dichloropropene	0.585	0.606	-3.6	81	0.00
55 T	1,1,2-Trichloroethane	0.438	0.458	-4.6	86	0.00
56 T	Ethyl methacrylate	0.607	0.665	-9.6	86	0.00
57 T	1,3-Dichloropropane	0.705	0.726	-3.0	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.293	0.325	-10.9	87	0.00
59 T	2-Hexanone	0.475	0.540	-13.7	90	0.00
60 T	Dibromochloromethane	0.457	0.495	-8.3	85	0.00
61 T	1,2-Dibromoethane	0.465	0.488	-4.9	85	0.00
62 S	4-Bromofluorobenzene	0.615	0.617	-0.3	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.592	0.647	-9.3	92	0.00
65 PM	Chlorobenzene	1.360	1.325	2.6	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.472	0.503	-6.6	86	0.00
67 C	Ethyl Benzene	2.241	2.176	2.9#	81	0.00
68 T	m/p-Xylenes	0.882	0.870	1.4	81	0.00
69 T	o-Xylene	0.855	0.849	0.7	83	0.00
70 T	Styrene	1.399	1.425	-1.9	84	0.00
71 P	Bromoform	0.399	0.421	-5.5	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.345	3.341	0.1	84	0.00
74 T	N-amyl acetate	1.435	1.477	-2.9	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.002	1.012	-1.0	85	0.00
76 T	1,2,3-Trichloropropane	0.925	0.987	-6.7	88	0.00
77 T	Bromobenzene	0.965	0.961	0.4	84	0.00
78 T	n-propylbenzene	3.717	3.741	-0.6	83	0.00
79 T	2-Chlorotoluene	2.243	2.214	1.3	82	0.00
80 T	1,3,5-Trimethylbenzene	2.809	2.853	-1.6	82	0.00
81 T	trans-1,4-Dichloro-2-butene	0.240	0.268	-11.7	85	0.00
82 T	4-Chlorotoluene	2.698	2.630	2.5	81	0.00
83 T	tert-Butylbenzene	2.801	2.821	-0.7	81	0.00
84 T	1,2,4-Trimethylbenzene	2.879	2.928	-1.7	82	0.00
85 T	sec-Butylbenzene	3.351	3.453	-3.0	84	0.00
86 T	p-Isopropyltoluene	3.074	3.179	-3.4	83	0.00
87 T	1,3-Dichlorobenzene	1.802	1.744	3.2	84	0.00
88 T	1,4-Dichlorobenzene	1.857	1.776	4.4	82	0.00
89 T	n-Butylbenzene	2.683	2.684	-0.0	81	0.00

Data Path : W:\HPCHEM1\MSVOA X\Data\VX050418\
Data File : VX001391.D
Acq On : 04 May 2018 11:40
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 04 14:28:10 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050118W.M
Quant Title : SW846 8260
QLast Update : Wed May 02 01:48:25 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.445	0.463	-4.0	81	0.00
91 T	1,2-Dichlorobenzene	1.793	1.789	0.2	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.203	0.239	-17.7	90	0.00
93 T	1,2,4-Trichlorobenzene	1.442	1.346	6.7	81	0.00
94 T	Hexachlorobutadiene	0.755	0.753	0.3	87	0.00
95 T	Naphthalene	3.659	3.910	-6.9	87	0.00
96 T	1,2,3-Trichlorobenzene	1.416	1.397	1.3	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6