


Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050518\
Data File : VX001412.D
Acq On : 05 May 2018 11:26
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Title : SW846 8260
Last Update : Sun May 06 07:59:25 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.2	6039	PASS
75	95	30	60	50.0	14907	PASS
95	95	100	100	100.0	29829	PASS
96	95	5	9	7.5	2239	PASS
173	174	0.00	2	1.1	328	PASS
174	95	50	100	97.0	28920	PASS
175	174	5	9	7.6	2209	PASS
176	174	95	101	97.5	28208	PASS
177	176	5	9	6.2	1739	PASS