


Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050618\
Data File : VX001436.D
Acq On : 06 May 2018 09:53
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Title : SW846 8260
Last Update : Sun May 06 07:59:25 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1723

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	7648	PASS
75	95	30	60	50.6	20213	PASS
95	95	100	100	100.0	39970	PASS
96	95	5	9	7.0	2799	PASS
173	174	0.00	2	1.1	417	PASS
174	95	50	100	96.3	38480	PASS
175	174	5	9	7.4	2843	PASS
176	174	95	101	97.7	37576	PASS
177	176	5	9	6.4	2392	PASS