

Data Path : W:\HPCHEM1\MSVOA X\Data\VX050718\
 Data File : VX001464.D
 Acq On : 07 May 2018 14:26
 Operator : JC/MD
 Sample : J2675-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-06-050418

Quant Time: May 08 06:50:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	271852	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	365886	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	319507	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	174809	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	210144	53.78	ug/l	0.00
Spiked Amount			Recovery	=	107.56%	
35) Dibromofluoromethane	5.51	113	165167	50.75	ug/l	0.00
Spiked Amount			Recovery	=	101.50%	
50) Toluene-d8	8.72	98	590332	49.69	ug/l	0.00
Spiked Amount			Recovery	=	99.38%	
62) 4-Bromofluorobenzene	11.14	95	193571	42.45	ug/l	0.00
Spiked Amount			Recovery	=	84.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	16323	9.20	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\Data\VX050718\
Data File : VX001464.D
Acq On : 07 May 2018 14:26
Operator : JC/MD
Sample : J2675-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HR-06-050418

Quant Time: May 08 06:50:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

