

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
 Data File : VX001465.D
 Acq On : 07 May 2018 14:53
 Operator : JC/MD
 Sample : J2659-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPS1-TT-MW303D-20180428

Quant Time: May 08 06:52:29 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	262261	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	356025	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	309280	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166553	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	213697	56.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.38%	
35) Dibromofluoromethane	5.51	113	166865	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
50) Toluene-d8	8.72	98	590955	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
62) 4-Bromofluorobenzene	11.14	95	191959	43.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.52%	
Target Compounds						
16) Acetone	2.45	43	9746	5.69	ug/l	94
44) Trichloroethene	7.22	130	7503	1.86	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
Data File : VX001465.D
Acq On : 07 May 2018 14:53
Operator : JC/MD
Sample : J2659-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPS1-TT-MW303D-20180428

Quant Time: May 08 06:52:29 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

