

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
 Data File : VX001485.D
 Acq On : 08 May 2018 00:40
 Operator : JC/MD
 Sample : J2680-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-02-050418-B

Quant Time: May 08 07:40:14 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251571	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	341068	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	302517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	164744	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	168928	46.72	ug/l	0.00
Spiked Amount			Recovery	=	93.44%	
35) Dibromofluoromethane	5.51	113	130440	43.00	ug/l	0.00
Spiked Amount			Recovery	=	86.00%	
50) Toluene-d8	8.72	98	470748	42.51	ug/l	0.00
Spiked Amount			Recovery	=	85.02%	
62) 4-Bromofluorobenzene	11.14	95	154726	36.40	ug/l	0.00
Spiked Amount			Recovery	=	72.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	4777	2.91	ug/l	96
18) Methyl Acetate	2.78	43	6088	1.41	ug/l	95
20) Methylene Chloride	2.86	84	67842	20.87	ug/l	97
44) Trichloroethene	7.22	130	10112	2.61	ug/l	93
68) m/p-Xylenes	10.36	106	24609	4.28	ug/l	98
69) o-Xylene	10.70	106	8895	1.59	ug/l	98
95) Naphthalene	13.84	128	16314	1.34	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
Data File : VX001485.D
Acq On : 08 May 2018 00:40
Operator : JC/MD
Sample : J2680-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-02-050418-B

Quant Time: May 08 07:40:14 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

