

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
 Data File : VX001488.D
 Acq On : 08 May 2018 02:00
 Operator : JC/MD
 Sample : J2680-10 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-02-050418-E

Quant Time: May 08 07:45:07 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	256795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	354403	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	311778	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	175278	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	171783	46.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.08%	
35) Dibromofluoromethane	5.51	113	133572	42.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.74%	
50) Toluene-d8	8.72	98	486871	42.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.62%	
62) 4-Bromofluorobenzene	11.14	95	161935	36.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.32%	
Target Compounds						
16) Acetone	2.45	43	4957	2.96	ug/l	99
18) Methyl Acetate	2.78	43	5767	1.31	ug/l	99
20) Methylene Chloride	2.86	84	7865	2.37	ug/l	96
44) Trichloroethene	7.22	130	8709	2.16	ug/l	91
95) Naphthalene	13.84	128	113405	8.73	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
Data File : VX001488.D
Acq On : 08 May 2018 02:00
Operator : JC/MD
Sample : J2680-10 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-02-050418-E

Quant Time: May 08 07:45:07 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

