

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
 Data File : VX001495.D
 Acq On : 08 May 2018 05:06
 Operator : JC/MD
 Sample : J2675-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-06-050418

Quant Time: May 08 08:01:25 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	262866	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	356915	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	320046	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	184572	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	176598	46.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.48%	
35) Dibromofluoromethane	5.51	113	137536	43.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.64%	
50) Toluene-d8	8.72	98	499147	43.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.14%	
62) 4-Bromofluorobenzene	11.14	95	166789	37.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	74.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	6571	3.83	ug/l	100
18) Methyl Acetate	2.78	43	6473	1.44	ug/l	95
20) Methylene Chloride	2.86	84	13963	4.11	ug/l	99
68) m/p-Xylenes	10.36	106	19166	3.15	ug/l	99
69) o-Xylene	10.70	106	12275	2.08	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	39009	3.67	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	152599	13.96	ug/l	100
95) Naphthalene	13.84	128	59847	4.38	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
Data File : VX001495.D
Acq On : 08 May 2018 05:06
Operator : JC/MD
Sample : J2675-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HR-06-050418

Quant Time: May 08 08:01:25 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

