

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\  
 Data File : VX001497.D  
 Acq On : 08 May 2018 05:59  
 Operator : JC/MD  
 Sample : J2738-02 5X  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 40 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 CB-TCLP-VAR2-0518

Quant Time: May 08 08:13:53 2018  
 Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X050518W.M  
 Quant Title : SW846 8260  
 QLast Update : Sun May 06 07:59:25 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|-------|--------|----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 265221   | 50.00 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.87  | 114  | 364841   | 50.00 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.12 | 117  | 326575   | 50.00 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.09 | 152  | 186295   | 50.00 | ug/l   | 0.00     |
| System Monitoring Compounds |       |      |          |       |        |          |
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 179167   | 47.00 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 94.00% |          |
| 35) Dibromofluoromethane    | 5.51  | 113  | 138819   | 42.78 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 85.56% |          |
| 50) Toluene-d8              | 8.72  | 98   | 503805   | 42.53 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 85.06% |          |
| 62) 4-Bromofluorobenzene    | 11.14 | 95   | 169834   | 37.35 | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =     | 74.70% |          |
| Target Compounds            |       |      |          |       |        |          |
| 16) Acetone                 | 2.45  | 43   | 6852     | 3.96  | ug/l   | 94       |
| 18) Methyl Acetate          | 2.78  | 43   | 5971     | 1.31  | ug/l   | 96       |
| 20) Methylene Chloride      | 2.86  | 84   | 5718     | 1.67  | ug/l   | 89       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\  
Data File : VX001497.D  
Acq On : 08 May 2018 05:59  
Operator : JC/MD  
Sample : J2738-02 5X  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 40 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
CB-TCLP-VAR2-0518

Quant Time: May 08 08:13:53 2018  
Quant Method : W:\HPCHEM1\MSVOA\_X\METHOD\82X050518W.M  
Quant Title : SW846 8260  
QLast Update : Sun May 06 07:59:25 2018  
Response via : Initial Calibration

