

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
 Data File : VX001501.D
 Acq On : 08 May 2018 07:45
 Operator : JC/MD
 Sample : J2736-03 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MS-SOIL

Quant Time: May 08 08:24:18 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	266486	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	366836	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	323979	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	183052	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	179447	46.85	ug/l	0.00
Spiked Amount			Recovery	=	93.70%	
35) Dibromofluoromethane	5.51	113	138738	42.52	ug/l	0.00
Spiked Amount			Recovery	=	85.04%	
50) Toluene-d8	8.72	98	507021	42.57	ug/l	0.00
Spiked Amount			Recovery	=	85.14%	
62) 4-Bromofluorobenzene	11.14	95	169839	37.15	ug/l	0.00
Spiked Amount			Recovery	=	74.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	4685	2.69	ug/l	95
18) Methyl Acetate	2.78	43	5916	1.29	ug/l	94
20) Methylene Chloride	2.86	84	20671	6.00	ug/l	96
67) Ethyl Benzene	10.26	91	45971	2.95	ug/l	99
68) m/p-Xylenes	10.36	106	13470	2.19	ug/l	100
69) o-Xylene	10.70	106	14967	2.50	ug/l	96
84) 1,2,4-Trimethylbenzene	11.81	105	34726	3.20	ug/l	98
95) Naphthalene	13.84	128	1015432	74.88	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
Data File : VX001501.D
Acq On : 08 May 2018 07:45
Operator : JC/MD
Sample : J2736-03 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MS-SOIL

Quant Time: May 08 08:24:18 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

