

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
 Data File : VX001502.D
 Acq On : 08 May 2018 08:12
 Operator : JC/MD
 Sample : J2677-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-03-050418

Quant Time: May 08 08:41:22 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264191	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	366212	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	323718	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	180912	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	179035	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
35) Dibromofluoromethane	5.51	113	138092	42.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.78%	
50) Toluene-d8	8.72	98	502839	42.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.58%	
62) 4-Bromofluorobenzene	11.14	95	168340	36.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.76%	
Target Compounds						
16) Acetone	2.45	43	5861	3.40	ug/l	98
18) Methyl Acetate	2.78	43	5608	1.24	ug/l	95
20) Methylene Chloride	2.86	84	232145	68.00	ug/l	97
95) Naphthalene	13.84	128	53913	4.02	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050718\
Data File : VX001502.D
Acq On : 08 May 2018 08:12
Operator : JC/MD
Sample : J2677-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OK-03-050418

Quant Time: May 08 08:41:22 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

