

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050818\
 Data File : VX001519.D
 Acq On : 08 May 2018 17:33
 Operator : JC/MD
 Sample : J2644-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0451-180425

Quant Time: May 10 11:21:57 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	231698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	319047	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	279075	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152184	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	195507	58.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.42%	
35) Dibromofluoromethane	5.51	113	153065	53.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.88%	
50) Toluene-d8	8.72	98	545009	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
62) 4-Bromofluorobenzene	11.15	95	180299	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1364	0.43	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	72242	23.71	ug/l	87
44) Trichloroethene	7.22	130	26595	7.34	ug/l	97
64) Tetrachloroethene	9.34	164	15774	4.06	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050818\
Data File : VX001519.D
Acq On : 08 May 2018 17:33
Operator : JC/MD
Sample : J2644-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0451-180425

Quant Time: May 10 11:21:57 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

