

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050818\
 Data File : VX001533.D
 Acq On : 09 May 2018 00:41
 Operator : JC/MD
 Sample : J2671-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-02-050718-A

Quant Time: May 09 07:50:03 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	242694	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	324056	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	285110	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152471	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161742	46.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.74%	
35) Dibromofluoromethane	5.51	113	126741	43.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.94%	
50) Toluene-d8	8.72	98	449183	42.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.38%	
62) 4-Bromofluorobenzene	11.14	95	145051	35.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.82%	
Target Compounds						
16) Acetone	2.45	43	5978	3.77	ug/l	91
18) Methyl Acetate	2.78	43	5809	1.40	ug/l	93
20) Methylene Chloride	2.86	84	17363	5.54	ug/l	98
44) Trichloroethene	7.22	130	18116	4.92	ug/l	94
68) m/p-Xylenes	10.36	106	10564	1.95	ug/l	97
95) Naphthalene	13.84	128	167024	14.79	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050818\
Data File : VX001533.D
Acq On : 09 May 2018 00:41
Operator : JC/MD
Sample : J2671-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-02-050718-A

Quant Time: May 09 07:50:03 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

