

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
 Data File : VX001555.D
 Acq On : 09 May 2018 14:22
 Operator : JC/MD
 Sample : VX0509WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0509WBL01

Quant Time: May 10 04:40:22 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250433	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	340085	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	298481	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	163276	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	207482	57.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.28%	
35) Dibromofluoromethane	5.51	113	164486	54.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.76%	
50) Toluene-d8	8.72	98	594810	53.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.74%	
62) 4-Bromofluorobenzene	11.14	95	193491	45.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.30%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
Data File : VX001555.D
Acq On : 09 May 2018 14:22
Operator : JC/MD
Sample : VX0509WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0509WBL01

Quant Time: May 10 04:40:22 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

