

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
 Data File : VX001556.D
 Acq On : 09 May 2018 15:46
 Operator : JC/MD
 Sample : J2732-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-05032018

Quant Time: May 10 04:42:35 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250130	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	342306	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	297938	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	156349	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	204946	57.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.02%	
35) Dibromofluoromethane	5.51	113	158516	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
50) Toluene-d8	8.72	98	566696	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
62) 4-Bromofluorobenzene	11.14	95	183173	42.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.86%	
Target Compounds						
16) Acetone	2.45	43	12791	7.84	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
Data File : VX001556.D
Acq On : 09 May 2018 15:46
Operator : JC/MD
Sample : J2732-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-05032018

Quant Time: May 10 04:42:35 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

