

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
 Data File : VX001568.D
 Acq On : 09 May 2018 21:07
 Operator : JC/MD
 Sample : J2732-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RAV-RMW-3

Quant Time: May 10 12:36:09 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226269	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	310260	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	267592	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	134050	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	190336	58.53	ug/l	0.00
Spiked Amount			Recovery	=	117.06%	
35) Dibromofluoromethane	5.51	113	150281	54.46	ug/l	0.00
Spiked Amount			Recovery	=	108.92%	
50) Toluene-d8	8.72	98	513992	51.02	ug/l	0.00
Spiked Amount			Recovery	=	102.04%	
62) 4-Bromofluorobenzene	11.14	95	159746	41.31	ug/l	0.00
Spiked Amount			Recovery	=	82.62%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
30) Chloroform	5.23	83	115086	22.93	ug/l	99
38) Carbon Tetrachloride	5.79	117	25468	6.05	ug/l	98
46) Dibromomethane	7.67	93	2736	1.34	ug/l	95
47) Bromodichloromethane	7.91	83	23579	6.26	ug/l	99
60) Dibromochloromethane	9.59	129	14131	4.54	ug/l	99
71) Bromoform	10.86	173	7748	6.23	ug/l #	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
Data File : VX001568.D
Acq On : 09 May 2018 21:07
Operator : JC/MD
Sample : J2732-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RAV-RMW-3

Quant Time: May 10 12:36:09 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

