

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
 Data File : VX001592.D
 Acq On : 10 May 2018 08:13
 Operator : JC/MD
 Sample : J2807-11 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CONCRETE-FOOTING

Quant Time: May 10 12:21:05 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	260833	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	357543	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	314784	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	176210	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	174574	46.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.14%	
35) Dibromofluoromethane	5.51	113	132985	41.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.64%	
50) Toluene-d8	8.72	98	491626	42.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.70%	
62) 4-Bromofluorobenzene	11.14	95	160620	36.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.08%	
Target Compounds						
68) m/p-Xylenes	10.36	106	3538	0.59	ug/l	98
69) o-Xylene	10.70	106	1408	0.24	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	6428	0.63	ug/l	100
84) 1,2,4-Trimethylbenzene	11.81	105	21483	2.06	ug/l	96
95) Naphthalene	13.84	128	7390	0.57	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
Data File : VX001592.D
Acq On : 10 May 2018 08:13
Operator : JC/MD
Sample : J2807-11 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CONCRETE-FOOTING

Quant Time: May 10 12:21:05 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

