

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
 Data File : VX001593.D
 Acq On : 10 May 2018 08:40
 Operator : JC/MD
 Sample : J2811-03 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-4A

Quant Time: May 10 11:31:24 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	262290	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	357322	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	318295	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	176013	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	177203	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
35) Dibromofluoromethane	5.51	113	138255	43.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.00%	
50) Toluene-d8	8.72	98	495778	42.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.46%	
62) 4-Bromofluorobenzene	11.14	95	162363	36.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.92%	
Target Compounds						
16) Acetone	2.45	43	5313	3.10	ug/l	Qvalue 100
18) Methyl Acetate	2.78	43	6937	1.54	ug/l	99
20) Methylene Chloride	2.86	84	14410	4.25	ug/l	96
44) Trichloroethene	7.22	130	6949	1.71	ug/l	93
68) m/p-Xylenes	10.36	106	6260	1.04	ug/l	97
69) o-Xylene	10.70	106	3287	0.56	ug/l	94
95) Naphthalene	13.84	128	100600	7.72	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
Data File : VX001593.D
Acq On : 10 May 2018 08:40
Operator : JC/MD
Sample : J2811-03 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-4A

Quant Time: May 10 11:31:24 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

