

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
 Data File : VX001595.D
 Acq On : 10 May 2018 09:34
 Operator : JC/MD
 Sample : J2811-09 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6B

Quant Time: May 10 11:38:41 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	256113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	347905	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	308233	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166606	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	172721	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
35) Dibromofluoromethane	5.51	113	133119	43.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.04%	
50) Toluene-d8	8.72	98	480693	42.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.12%	
62) 4-Bromofluorobenzene	11.14	95	155296	35.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.62%	
Target Compounds						
16) Acetone	2.45	43	6575	3.93	ug/l	99
18) Methyl Acetate	2.78	43	6019	1.37	ug/l	100
20) Methylene Chloride	2.86	84	23410	7.07	ug/l	97
95) Naphthalene	13.84	128	21830	1.77	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX050918\
Data File : VX001595.D
Acq On : 10 May 2018 09:34
Operator : JC/MD
Sample : J2811-09 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6B

Quant Time: May 10 11:38:41 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

