

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051018\
 Data File : VX001604.D
 Acq On : 10 May 2018 16:20
 Operator : JC/MD
 Sample : J2738-01 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-TCLP-VAR1-0518

Quant Time: May 11 07:43:21 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	245579	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	330778	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	288446	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	156720	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	198864	56.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.68%	
35) Dibromofluoromethane	5.51	113	154682	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
50) Toluene-d8	8.72	98	537505	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	11.14	95	177767	43.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.24%	
Target Compounds						
16) Acetone	2.45	43	8489	5.30	ug/l	91
18) Methyl Acetate	2.78	43	7560	1.80	ug/l	94
20) Methylene Chloride	2.86	84	63716	20.08	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051018\
Data File : VX001604.D
Acq On : 10 May 2018 16:20
Operator : JC/MD
Sample : J2738-01 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-TCLP-VAR1-0518

Quant Time: May 11 07:43:21 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

