

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051018\
 Data File : VX001607.D
 Acq On : 10 May 2018 17:40
 Operator : JC/MD
 Sample : J2738-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-TCLP-VAR4-0518

Quant Time: May 11 07:36:15 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	230545	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	313267	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	276644	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	150668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	191399	57.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.52%	
35) Dibromofluoromethane	5.51	113	151476	54.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.72%	
50) Toluene-d8	8.72	98	526566	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
62) 4-Bromofluorobenzene	11.14	95	172948	44.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.58%	
Target Compounds						
16) Acetone	2.45	43	5967	3.97	ug/l	94
18) Methyl Acetate	2.79	43	6911	1.75	ug/l	100
20) Methylene Chloride	2.85	84	4775	1.60	ug/l	88
44) Trichloroethene	7.22	130	10181	2.86	ug/l	93
68) m/p-Xylenes	10.36	106	6132	1.17	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	24857	2.79	ug/l	99
95) Naphthalene	13.84	128	90435	8.10	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051018\
Data File : VX001607.D
Acq On : 10 May 2018 17:40
Operator : JC/MD
Sample : J2738-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-TCLP-VAR4-0518

Quant Time: May 11 07:36:15 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

