

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX051118\
 Data File : VX001628.D
 Acq On : 11 May 2018 03:55
 Operator : JC/MD
 Sample : J2766-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(MAY)

Quant Time: May 11 07:24:48 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\624X051118W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri May 11 07:16:13 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	54778	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	286536	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	252890	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	120832	30.23	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.77%
60) 4-Bromofluorobenzene	11.14	95	118022	29.49	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	98.30%
63) Toluene-d8	8.72	98	334406	29.65	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.83%

Target Compounds


						Qvalue
5) Bromomethane	1.65	94	1003	0.66	ug/l	98
6) Chloroethane	1.32	64	721727	351.29	ug/l #	54
11) Methyl Iodide	2.51	142	2448	0.66	ug/l	90
12) Methyl Acetate	2.78	43	21217	4.92	ug/l	98
13) Acrolein	2.30	56	387	0.79	ug/l	99
15) Acetone	2.45	58	1098709	2374.96	ug/l	93
16) Carbon Disulfide	2.57	76	9684	1.54	ug/l	100
17) Allyl chloride	2.63	41	34881	7.21	ug/l	76
18) Methylene Chloride	2.86	84	2165	0.72	ug/l #	83
23) tert-Butyl Alcohol	3.09	59	504	0.66	ug/l #	87
30) 2-Butanone	4.72	43	54686	23.02	ug/l	100
45) 1,4-Dioxane	7.78	88	68	0.84	ug/l #	1
53) Dibromochloromethane	9.34	129	6930	2.18	ug/l #	12
56) Bromoform	11.15	173	1495	0.58	ug/l #	1
58) 4-Methyl-2-Pentanone	8.66	43	15013	3.30	ug/l	97
59) 2-Hexanone	9.50	43	1276	0.37	ug/l #	90
61) Tetrachloroethene	9.34	164	7669	2.08	ug/l	96
62) Toluene	8.79	91	42046	3.46	ug/l	98
66) Ethyl Benzene	10.26	91	5133	0.41	ug/l	96
67) m/p-Xylenes	10.36	106	2401	0.48	ug/l	99
68) o-Xylene	10.71	106	1159	0.23	ug/l	97
69) Styrene	10.72	104	2699	0.33	ug/l	98
77) t-1,4-Dichloro-2-butene	11.14	75	58879	54.83	ug/l #	4
79) tert-butylbenzene	12.07	119	240829	21.05	ug/l	99
80) 1,2,4-Trimethylbenzene	11.82	105	17022	1.54	ug/l	54
82) p-Isopropyltoluene	12.07	119	240829	21.05	ug/l	99
85) n-Butylbenzene	12.30	91	13865	1.52	ug/l	76
86) Hexachloroethane	12.59	117	1179	0.65	ug/l #	1
91) Naphthalene	13.84	128	7542	0.50	ug/l #	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX051118\
Data File : VX001628.D
Acq On : 11 May 2018 03:55
Operator : JC/MD
Sample : J2766-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
002M-35TH-AVE(MAY)

Quant Time: May 11 07:24:48 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\624X051118W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri May 11 07:16:13 2018
Response via : Initial Calibration









#1
 Bromochloromethane
 Concen: 30.00 ug/l
 RT: 5.03 min Scan# 728
 Delta R.T. 0.00 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55




Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(MAY)


Tgt Ion:128 Resp: 54778
 Ion Ratio Lower Upper
 128 100
 49 146.3 0.0 372.0
 130 129.7 0.0 326.5



#5
 Bromomethane
 Concen: 0.66 ug/l
 RT: 1.65 min Scan# 174
 Delta R.T. 0.01 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55

Tgt Ion: 94 Resp: 1003
 Ion Ratio Lower Upper
 94 100
 96 95.0 77.8 116.8





#6

Chloroethane

Concen: 351.29 ug/l

RT: 1.32 min Scan# 121

Delta R.T. -0.40 min

Lab File: VX001628.D

Acq: 11 May 2018 03:55

Instrument :

MSVOA_X

ClientSampleId :


002M-35TH-AVE(MAY)

Tgt Ion: 64 Resp: 721727

Ion Ratio Lower Upper




64 100

66 4.6 23.4 35.2#



Abundance Ion 64.00 (63.70 to 64.70): VX0

Ion 66.00 (65.70 to 66.70): VX0



#11

Methyl Iodide

Concen: 0.66 ug/l

RT: 2.51 min Scan# 316

Delta R.T. 0.00 min

Lab File: VX001628.D

Acq: 11 May 2018 03:55


Tgt Ion: 142 Resp: 2448

Ion Ratio Lower Upper

142 100

127 48.8 20.7 62.1



141 13.1 7.4 22.3




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V





#12

Methyl Acetate

Concen: 4.92 ug/l

RT: 2.78 min Scan# 360

Delta R.T. 0.00 min

Lab File: VX001628.D

Acq: 11 May 2018 03:55

Instrument :

MSVOA_X

ClientSampleId :





002M-35TH-AVE(MAY)

Tgt Ion: 43 Resp: 21217

Ion Ratio Lower Upper

43 100

74 21.3 18.0 27.0



#13

Acrolein

Concen: 0.79 ug/l

RT: 2.30 min Scan# 281

Delta R.T. 0.00 min

Lab File: VX001628.D




Acq: 11 May 2018 03:55


Tgt Ion: 56 Resp: 387

Ion Ratio Lower Upper

56 100

55 71.8 56.9 85.3





#15

Acetone

Concen: 2374.96 ug/l

RT: 2.45 min Scan# 306

Delta R.T. 0.00 min

Lab File: VX001628.D

Acq: 11 May 2018 03:55

Instrument :

MSVOA_X

ClientSampleId :


002M-35TH-AVE(MAY)

Tgt Ion: 58 Resp: 1098709

Ion Ratio Lower Upper

58 100

43 316.0 264.2 396.2



Abundance Ion 58.00 (57.70 to 58.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

2500000

2000000

1500000



1000000

500000

0

Time-->

2.40 2.45 2.50 2.60



#16

Carbon Disulfide

Concen: 1.54 ug/l

RT: 2.57 min Scan# 326

Delta R.T. 0.01 min

Lab File: VX001628.D


Acq: 11 May 2018 03:55

Tgt Ion: 76 Resp: 9684

Ion Ratio Lower Upper

76 100

78 9.2 7.3 10.9



Abundance Ion 76.00 (75.70 to 76.70): VX0

Ion 78.00 (77.70 to 78.70): VX0

8000

6000


4000


2000

0

Time-->

2.55 2.57 2.60









#17
 Allyl chloride
 Concen: 7.21 ug/l
 RT: 2.63 min Scan# 335
 Delta R.T. -0.10 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55




Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(MAY)


Tgt Ion: 41 Resp: 34881
 Ion Ratio Lower Upper
 41 100
 39 87.2 33.7 101.1
 76 19.2 15.4 46.4



#18
 Methylene Chloride
 Concen: 0.72 ug/l
 RT: 2.86 min Scan# 373
 Delta R.T. 0.00 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55

Tgt Ion: 84 Resp: 2165
 Ion Ratio Lower Upper
 84 100
 49 106.9 99.5 149.3
 51 49.0 30.1 45.1#
 86 53.6 55.4 83.2#









#23
 tert-Butyl Alcohol
 Concen: 0.66 ug/l
 RT: 3.09 min Scan# 410
 Delta R.T. 0.01 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55




Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(MAY)


Tgt Ion: 59 Resp: 504
 Ion Ratio Lower Upper
 59 100
 52 16.9 0.2 0.2#



#27
 1,2-Dichloroethane-d4
 Concen: 30.23 ug/l
 RT: 6.07 min Scan# 900
 Delta R.T. -0.01 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55

Tgt Ion: 65 Resp: 120832
 Ion Ratio Lower Upper
 65 100
 67 50.8 40.2 60.2









#28
 1,4-Difluorobenzene
 Concen: 30.00 ug/l
 RT: 6.87 min Scan# 1030
 Delta R.T. 0.00 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55




Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(MAY)


Tgt Ion: 114 Resp: 286536
 Ion Ratio Lower Upper
 114 100
 63 19.3 15.5 23.3
 88 15.0 12.2 18.4



#30
 2-Butanone
 Concen: 23.02 ug/l
 RT: 4.72 min Scan# 678
 Delta R.T. 0.00 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55

Tgt Ion: 43 Resp: 54686
 Ion Ratio Lower Upper
 43 100
 57 7.9 6.3 9.5
 72 25.1 20.2 30.4









#45
 1,4-Dioxane
 Concen: 0.84 ug/l
 RT: 7.78 min Scan# 1180
 Delta R.T. 0.02 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55




Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(MAY)


Tgt Ion: 88 Resp: 68
 Ion Ratio Lower Upper
 88 100
 43 269.1 27.9 41.9#
 58 64.7 55.3 82.9



#53
 Dibromochloromethane
 Concen: 2.18 ug/l
 RT: 9.34 min Scan# 1436
 Delta R.T. -0.25 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55

Tgt Ion: 129 Resp: 6930
 Ion Ratio Lower Upper
 129 100
 127 0.0 37.9 113.7#





#56

Bromoform

Concen: 0.58 ug/l

RT: 11.15 min Scan# 1733

Delta R.T. 0.29 min

Lab File: VX001628.D

Acq: 11 May 2018 03:55

Instrument :

MSVOA_X

ClientSampleId :

002M-35TH-AVE(MAY)


Tgt Ion:173 Resp: 1495

Ion Ratio Lower Upper

173 100

175 590.9 39.8 59.8#




254 0.0 10.7 16.1#



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#57

Chlorobenzene-d5

Concen: 30.00 ug/l

RT: 10.12 min Scan# 1563

Delta R.T. -0.01 min

Lab File: VX001628.D

Acq: 11 May 2018 03:55


Tgt Ion:117 Resp: 252890

Ion Ratio Lower Upper

117 100

82 52.3 42.2 63.2



119 32.2 26.3 39.5




Abundance Ion 117.00 (116.70 to 117.70): V

Ion 82.00 (81.70 to 82.70): V

Ion 119.00 (118.70 to 119.70): V









#58
 4-Methyl-2-Pentanone
 Concen: 3.30 ug/l
 RT: 8.66 min Scan# 1324
 Delta R.T. 0.00 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55

Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(MAY)

Tgt Ion: 43 Resp: 15013
 Ion Ratio Lower Upper
 43 100
 58 39.7 30.4 45.6
 85 16.4 14.3 21.5




Abundance Ion 43.00 (42.70 to 43.70): VX0
 Ion 58.00 (57.70 to 58.70): VX0
 Ion 85.00 (84.70 to 85.70): VX0






#59
 2-Hexanone
 Concen: 0.37 ug/l
 RT: 9.50 min Scan# 1462
 Delta R.T. -0.01 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55

Tgt Ion: 43 Resp: 1276
 Ion Ratio Lower Upper
 43 100
 58 46.8 43.0 64.4
 57 23.4 15.4 23.0#



Abundance Ion 43.00 (42.70 to 43.70): VX0
 Ion 58.00 (57.70 to 58.70): VX0
 Ion 57.00 (56.70 to 57.70): VX0









#60
4-Bromofluorobenzene
Concen: 29.49 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. 0.00 min
Lab File: VX001628.D
Acq: 11 May 2018 03:55




Instrument :
MSVOA_X
ClientSampleId :
002M-35TH-AVE(MAY)


Tgt Ion: 95 Resp: 118022
Ion Ratio Lower Upper
95 100
174 100.6 80.2 120.2
176 98.8 79.2 118.8



#61
Tetrachloroethene
Concen: 2.08 ug/l
RT: 9.34 min Scan# 1436
Delta R.T. 0.00 min
Lab File: VX001628.D
Acq: 11 May 2018 03:55

Tgt Ion: 164 Resp: 7669
Ion Ratio Lower Upper
164 100
166 135.3 102.8 154.2
129 83.4 67.8 101.8
131 84.9 66.2 99.2





#62

Toluene

Concen: 3.46 ug/l

RT: 8.79 min Scan# 1346

Delta R.T. 0.00 min

Lab File: VX001628.D

Acq: 11 May 2018 03:55

Instrument :

MSVOA_X

ClientSampleId :





002M-35TH-AVE(MAY)

Tgt Ion: 91 Resp: 42046

Ion Ratio Lower Upper

91 100

92 57.7 47.4 71.0



#63

Toluene-d8

Concen: 29.65 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX001628.D




Acq: 11 May 2018 03:55


Tgt Ion: 98 Resp: 334406

Ion Ratio Lower Upper

98 100

100 66.2 52.2 78.2









#66
Ethyl Benzene
Concen: 0.41 ug/l
RT: 10.26 min Scan# 1586
Delta R.T. 0.00 min
Lab File: VX001628.D
Acq: 11 May 2018 03:55




Instrument :
MSVOA_X
ClientSampleId :
002M-35TH-AVE(MAY)


Tgt Ion: 91 Resp: 5133
Ion Ratio Lower Upper
91 100
106 35.0 26.2 39.4



#67
m/p-Xylenes
Concen: 0.48 ug/l
RT: 10.36 min Scan# 1603
Delta R.T. -0.01 min
Lab File: VX001628.D
Acq: 11 May 2018 03:55

Tgt Ion: 106 Resp: 2401
Ion Ratio Lower Upper
106 100
91 195.1 157.8 236.8









#68
o-Xylene
Concen: 0.23 ug/l
RT: 10.71 min Scan# 1660
Delta R.T. 0.00 min
Lab File: VX001628.D
Acq: 11 May 2018 03:55

Instrument :
MSVOA_X
ClientSampleId :
002M-35TH-AVE(MAY)

Tgt Ion:106 Resp: 1159
Ion Ratio Lower Upper
106 100
91 204.4 104.3 312.8




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0






#69
Styrene
Concen: 0.33 ug/l
RT: 10.72 min Scan# 1662
Delta R.T. 0.00 min
Lab File: VX001628.D
Acq: 11 May 2018 03:55

Tgt Ion:104 Resp: 2699
Ion Ratio Lower Upper
104 100
78 51.1 39.3 58.9
103 53.4 43.7 65.5



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VX0
Ion 103.00 (102.70 to 103.70): V









#77
 t-1,4-Dichloro-2-butene
 Concen: 54.83 ug/l
 RT: 11.14 min Scan# 1731
 Delta R.T. 0.06 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55




Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(MAY)


Tgt Ion: 75 Resp: 58879
 Ion Ratio Lower Upper
 75 100
 53 0.2 58.3 174.8#
 89 0.0 23.9 71.8#



#79
 tert-butylbenzene
 Concen: 21.05 ug/l
 RT: 12.07 min Scan# 1883
 Delta R.T. -0.01 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55

Tgt Ion: 119 Resp: 240829
 Ion Ratio Lower Upper
 119 100
 91 22.4 0.0 44.0
 134 26.7 0.0 53.8









#80
 1,2,4-Trimethylbenzene
 Concen: 1.54 ug/l
 RT: 11.82 min Scan# 1842
 Delta R.T. 0.01 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55




Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(MAY)


Tgt Ion:105 Resp: 17022
 Ion Ratio Lower Upper
 105 100
 120 16.4 0.0 94.0



#82
 p-Isopropyltoluene
 Concen: 21.05 ug/l
 RT: 12.07 min Scan# 1883
 Delta R.T. -0.01 min
 Lab File: VX001628.D
 Acq: 11 May 2018 03:55

Tgt Ion:119 Resp: 240829
 Ion Ratio Lower Upper
 119 100
 134 26.7 0.0 53.8
 91 22.4 0.0 44.0





#85

n-Butylbenzene

Concen: 1.52 ug/l

RT: 12.30 min Scan# 1921

Delta R.T. -0.10 min

Lab File: VX001628.D

Acq: 11 May 2018 03:55

Instrument :

MSVOA_X

ClientSampleId :

002M-35TH-AVE(MAY)





Tgt Ion: 91 Resp: 13865

Ion Ratio Lower Upper

91 100

92 38.2 0.0 107.8

134 14.2 0.0 56.8



#86

Hexachloroethane

Concen: 0.65 ug/l

RT: 12.59 min Scan# 1968

Delta R.T. -0.02 min

Lab File: VX001628.D




Acq: 11 May 2018 03:55


Tgt Ion: 117 Resp: 1179

Ion Ratio Lower Upper

117 100

201 0.0 50.1 150.3#





#91
Naphthalene
Concen: 0.50 ug/l
RT: 13.84 min Scan# 2174
Delta R.T. 0.00 min
Lab File: VX001628.D
Acq: 11 May 2018 03:55

Instrument :
MSVOA_X
ClientSampleId :
002M-35TH-AVE(MAY)

Tgt Ion:128 Resp: 7542
Ion Ratio Lower Upper
128 100
127 14.7 10.3 15.5
129 14.3 8.8 13.2#

