

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
 Data File : VX001641.D
 Acq On : 11 May 2018 16:07
 Operator : JC/MD
 Sample : J2741-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW309-180501

Quant Time: May 12 12:08:06 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254793	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	347349	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	304785	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	163527	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	199358	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	5.51	113	159514	51.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.26%	
50) Toluene-d8	8.72	98	560159	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
62) 4-Bromofluorobenzene	11.15	95	184175	42.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.08%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	2317	0.74	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	290635	86.73	ug/l	86
44) Trichloroethene	7.22	130	99618	25.25	ug/l	99
64) Tetrachloroethene	9.34	164	1017520	239.51	ug/l	100
65) Chlorobenzene	10.14	112	3396	0.38	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
Data File : VX001641.D
Acq On : 11 May 2018 16:07
Operator : JC/MD
Sample : J2741-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW309-180501

Quant Time: May 12 12:08:06 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

