

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
 Data File : VX001643.D
 Acq On : 11 May 2018 17:01
 Operator : JC/MD
 Sample : J2741-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW151-180501

Quant Time: May 12 07:00:57 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	256733	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	343991	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	299733	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	161703	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	207539	56.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.48%	
35) Dibromofluoromethane	5.51	113	165254	54.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.02%	
50) Toluene-d8	8.72	98	581875	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
62) 4-Bromofluorobenzene	11.15	95	190419	44.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.82%	
Target Compounds						
4) Vinyl Chloride	1.40	62	35615	10.14	ug/l	95
12) 1,1-Dichloroethene	2.38	96	1992	0.70	ug/l #	87
21) trans-1,2-Dichloroethene	3.17	96	1918	0.60	ug/l #	81
27) cis-1,2-Dichloroethene	4.60	96	320326	94.87	ug/l	87
44) Trichloroethene	7.22	130	20602	5.27	ug/l	99
64) Tetrachloroethene	9.34	164	18267	4.37	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
Data File : VX001643.D
Acq On : 11 May 2018 17:01
Operator : JC/MD
Sample : J2741-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW151-180501

Quant Time: May 12 07:00:57 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

