

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
 Data File : VX001646.D
 Acq On : 11 May 2018 18:21
 Operator : JC/MD
 Sample : J2741-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0453-180501

Quant Time: May 12 07:06:34 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	242738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	331157	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	291739	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	158511	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	199622	57.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.44%	
35) Dibromofluoromethane	5.51	113	158659	53.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.72%	
50) Toluene-d8	8.72	98	557038	51.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.62%	
62) 4-Bromofluorobenzene	11.15	95	184159	44.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.24%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	2353	0.78	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	279255	87.47	ug/l	87
44) Trichloroethene	7.22	130	95475	25.39	ug/l	100
64) Tetrachloroethene	9.34	164	1013923	249.34	ug/l	99
65) Chlorobenzene	10.15	112	4077	0.48	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
Data File : VX001646.D
Acq On : 11 May 2018 18:21
Operator : JC/MD
Sample : J2741-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0453-180501

Quant Time: May 12 07:06:34 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

