

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
 Data File : VX001647.D
 Acq On : 11 May 2018 18:48
 Operator : JC/MD
 Sample : J2741-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW301-180502

Quant Time: May 12 07:08:13 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252624	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	337455	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	294475	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	157377	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	203066	55.93	ug/l	0.00
Spiked Amount			Recovery	=	111.86%	
35) Dibromofluoromethane	5.51	113	163022	54.31	ug/l	0.00
Spiked Amount			Recovery	=	108.62%	
50) Toluene-d8	8.72	98	566328	51.69	ug/l	0.00
Spiked Amount			Recovery	=	103.38%	
62) 4-Bromofluorobenzene	11.14	95	184891	43.96	ug/l	0.00
Spiked Amount			Recovery	=	87.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	252369	73.01	ug/l	99
12) 1,1-Dichloroethene	2.37	96	5680	2.03	ug/l	95
21) trans-1,2-Dichloroethene	3.17	96	5824	1.87	ug/l	99
24) 1,1-Dichloroethane	3.70	63	9002	1.65	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	2607942	784.94	ug/l	85
44) Trichloroethene	7.22	130	4116	1.07	ug/l	99
64) Tetrachloroethene	9.34	164	8024	1.95	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
Data File : VX001647.D
Acq On : 11 May 2018 18:48
Operator : JC/MD
Sample : J2741-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW301-180502

Quant Time: May 12 07:08:13 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

