

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
 Data File : VX001650.D
 Acq On : 11 May 2018 20:08
 Operator : JC/MD
 Sample : J2741-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW008-180430

Quant Time: May 12 07:17:17 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	235080	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	316187	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	270525	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	132733	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	195162	57.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.52%	
35) Dibromofluoromethane	5.51	113	155697	55.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.72%	
50) Toluene-d8	8.72	98	531377	51.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.52%	
62) 4-Bromofluorobenzene	11.14	95	165215	41.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.84%	
Target Compounds						
4) Vinyl Chloride	1.40	62	7257	2.26	ug/l	# 68
27) cis-1,2-Dichloroethene	4.60	96	22584	7.30	ug/l	90
44) Trichloroethene	7.22	130	3117	0.87	ug/l	83
64) Tetrachloroethene	9.34	164	13792	3.66	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
Data File : VX001650.D
Acq On : 11 May 2018 20:08
Operator : JC/MD
Sample : J2741-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW008-180430

Quant Time: May 12 07:17:17 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

