

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
 Data File : VX001652.D
 Acq On : 11 May 2018 21:02
 Operator : JC/MD
 Sample : J2741-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M-14-180502

Quant Time: May 12 07:22:06 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248155	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	332631	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	297121	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	158410	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	200128	56.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.22%	
35) Dibromofluoromethane	5.51	113	159821	54.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.04%	
50) Toluene-d8	8.72	98	556172	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
62) 4-Bromofluorobenzene	11.14	95	182940	44.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.26%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	2175	0.71	ug/l	84
27) cis-1,2-Dichloroethene	4.60	96	126480	38.75	ug/l	86
30) Chloroform	5.23	83	3265	0.59	ug/l	98
44) Trichloroethene	7.22	130	64664	17.12	ug/l	95
64) Tetrachloroethene	9.34	164	461492	111.43	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
Data File : VX001652.D
Acq On : 11 May 2018 21:02
Operator : JC/MD
Sample : J2741-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M-14-180502

Quant Time: May 12 07:22:06 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

