

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
 Data File : VX001653.D
 Acq On : 11 May 2018 21:28
 Operator : JC/MD
 Sample : J2741-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW012-180502

Quant Time: May 12 07:24:16 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	239497	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	324707	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	283773	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	150547	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	240296	69.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	139.62%	
35) Dibromofluoromethane	5.51	113	191229	66.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	132.42%	
50) Toluene-d8	8.72	98	663853	62.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.94%	
62) 4-Bromofluorobenzene	11.14	95	216985	53.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.24%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	14734	4.68	ug/l	96
30) Chloroform	5.23	83	2081	0.39	ug/l	98
44) Trichloroethene	7.22	130	9631	2.61	ug/l	97
64) Tetrachloroethene	9.34	164	46567	11.77	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051118\
Data File : VX001653.D
Acq On : 11 May 2018 21:28
Operator : JC/MD
Sample : J2741-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW012-180502

Quant Time: May 12 07:24:16 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

