

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001660.D
 Acq On : 12 May 2018 12:08
 Operator : JC/MD
 Sample : VX0512WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0512WBL01

Quant Time: May 14 06:37:53 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	285190	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	383658	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	319611	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	156709	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	234909	57.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.62%	
35) Dibromofluoromethane	5.51	113	188883	55.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.70%	
50) Toluene-d8	8.72	98	666229	53.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.96%	
62) 4-Bromofluorobenzene	11.14	95	204564	42.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.56%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001660.D
Acq On : 12 May 2018 12:08
Operator : JC/MD
Sample : VX0512WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0512WBL01

Quant Time: May 14 06:37:53 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

