

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001663.D
 Acq On : 12 May 2018 13:36
 Operator : JC/MD
 Sample : J2741-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW167-180502

Quant Time: May 14 06:44:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	276135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	370770	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	327682	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	176838	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	220273	55.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.00%	
35) Dibromofluoromethane	5.51	113	175963	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
50) Toluene-d8	8.72	98	626301	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
62) 4-Bromofluorobenzene	11.14	95	206316	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
Target Compounds						
21) trans-1,2-Dichloroethene	3.17	96	2221	0.65	ug/l	88
27) cis-1,2-Dichloroethene	4.60	96	685991	188.89	ug/l	85
44) Trichloroethene	7.22	130	54596	12.97	ug/l	99
64) Tetrachloroethene	9.34	164	1277031	279.60	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001663.D
Acq On : 12 May 2018 13:36
Operator : JC/MD
Sample : J2741-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW167-180502

Quant Time: May 14 06:44:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

