

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001676.D
 Acq On : 12 May 2018 19:22
 Operator : JC/MD
 Sample : J2885-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4TB01-180508

Quant Time: May 14 13:51:07 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	261836	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	353442	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	309732	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	162068	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	209697	55.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.44%	
35) Dibromofluoromethane	5.51	113	169644	53.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.92%	
50) Toluene-d8	8.72	98	592931	51.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.34%	
62) 4-Bromofluorobenzene	11.14	95	187584	42.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.16%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001676.D
Acq On : 12 May 2018 19:22
Operator : JC/MD
Sample : J2885-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4TB01-180508

Quant Time: May 14 13:51:07 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

