

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001690.D
 Acq On : 13 May 2018 02:01
 Operator : JC/MD
 Sample : PB109200ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109200ZHE#06

Quant Time: May 14 07:47:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	270035	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	368846	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	321216	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	174991	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	177642	45.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.54%	
35) Dibromofluoromethane	5.51	113	143477	43.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.46%	
50) Toluene-d8	8.72	98	507958	42.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.84%	
62) 4-Bromofluorobenzene	11.14	95	160602	34.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.86%	
Target Compounds						
16) Acetone	2.45	43	7018	3.98	ug/l	98
18) Methyl Acetate	2.78	43	23545	5.08	ug/l	100
20) Methylene Chloride	2.87	84	2475	0.71	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001690.D
Acq On : 13 May 2018 02:01
Operator : JC/MD
Sample : PB109200ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109200ZHE#06

Quant Time: May 14 07:47:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

