

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001701.D
 Acq On : 13 May 2018 06:52
 Operator : JC/MD
 Sample : PB109200ZHE#17
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109200ZHE#17

Quant Time: May 14 08:00:59 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	278029	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	378800	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	332628	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	180378	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	183876	46.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.02%	
35) Dibromofluoromethane	5.51	113	147046	43.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.28%	
50) Toluene-d8	8.72	98	530099	43.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.20%	
62) 4-Bromofluorobenzene	11.14	95	165083	34.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.94%	
Target Compounds						
16) Acetone	2.45	43	7666	4.22	ug/l	99
18) Methyl Acetate	2.78	43	22878	4.80	ug/l	99
20) Methylene Chloride	2.86	84	3171	0.88	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001701.D
Acq On : 13 May 2018 06:52
Operator : JC/MD
Sample : PB109200ZHE#17
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109200ZHE#17

Quant Time: May 14 08:00:59 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

