

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001702.D
 Acq On : 13 May 2018 07:19
 Operator : JC/MD
 Sample : PB109200ZHE#18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109200ZHE#18

Quant Time: May 14 08:01:43 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	272580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	372211	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	327960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	176441	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	180632	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
35) Dibromofluoromethane	5.51	113	145664	44.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.00%	
50) Toluene-d8	8.72	98	521991	43.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.38%	
62) 4-Bromofluorobenzene	11.14	95	164078	35.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.74%	
Target Compounds						
16) Acetone	2.45	43	7541	4.24	ug/l	99
18) Methyl Acetate	2.78	43	23006	4.92	ug/l	100
20) Methylene Chloride	2.86	84	2413	0.69	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001702.D
Acq On : 13 May 2018 07:19
Operator : JC/MD
Sample : PB109200ZHE#18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109200ZHE#18

Quant Time: May 14 08:01:43 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

