

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001703.D
 Acq On : 13 May 2018 07:45
 Operator : JC/MD
 Sample : PB109200ZHE#19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109200ZHE#19

Quant Time: May 14 08:03:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277608	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	372549	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	331230	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	179212	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	180955	45.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.70%	
35) Dibromofluoromethane	5.51	113	145739	43.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.96%	
50) Toluene-d8	8.72	98	524117	43.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.66%	
62) 4-Bromofluorobenzene	11.14	95	165430	35.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.26%	
Target Compounds						
16) Acetone	2.45	43	7348	4.06	ug/l	94
18) Methyl Acetate	2.78	43	22706	4.77	ug/l	98
20) Methylene Chloride	2.86	84	2675	0.75	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001703.D
Acq On : 13 May 2018 07:45
Operator : JC/MD
Sample : PB109200ZHE#19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109200ZHE#19

Quant Time: May 14 08:03:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

