

Data Path : W:\HPCHEM1\MSVOA X\DATA\X051218\
 Data File : VX001678.D
 Acq On : 12 May 2018 20:16
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 14 07:24:19 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.673	0.599	11.0	76	0.00
3 P	Chloromethane	0.656	0.588	10.4	81	0.00
4 C	Vinyl Chloride	0.684	0.629	8.0#	82	0.00
5 T	Bromomethane	0.400	0.244	39.0#	68	0.00
6 T	Chloroethane	0.452	0.425	6.0	85	0.00
7 T	Trichlorofluoromethane	1.030	0.979	5.0	84	0.00
8 T	Diethyl Ether	0.404	0.391	3.2	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.604	0.581	3.8	86	0.00
10 T	Methyl Iodide	0.871	0.689	20.9#	67	0.00
11 T	Tert butyl alcohol	0.152	0.147	3.3	86	0.00
12 CM	1,1-Dichloroethene	0.553	0.524	5.2#	84	0.00
13 T	Acrolein	0.035	0.019	45.7#	57	0.00
14 T	Allyl chloride	1.032	1.018	1.4	87	0.00
15 T	Acrylonitrile	0.342	0.338	1.2	88	0.00
16 T	Acetone	0.326	0.297	8.9	84	0.00
17 T	Carbon Disulfide	1.408	1.200	14.8	74	0.00
18 T	Methyl Acetate	0.857	0.949	-10.7	100	0.00
19 T	Methyl tert-butyl Ether	1.926	1.911	0.8	89	0.00
20 T	Methylene Chloride	0.646	0.612	5.3	88	0.00
21 T	trans-1,2-Dichloroethene	0.617	0.564	8.6	84	0.00
22 T	Diisopropyl ether	1.927	1.968	-2.1	92	0.00
23 T	Vinyl Acetate	1.606	1.603	0.2	87	0.00
24 P	1,1-Dichloroethane	1.077	1.046	2.9	87	0.00
25 T	2-Butanone	0.472	0.460	2.5	87	0.00
26 T	2,2-Dichloropropane	0.831	0.771	7.2	81	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.655	0.5	90	0.00
28 T	Bromochloromethane	0.467	0.491	-5.1	92	0.00
29 T	Tetrahydrofuran	0.302	0.296	2.0	88	0.00
30 C	Chloroform	1.109	1.162	-4.8#	93	0.00
31 T	Cyclohexane	0.925	0.883	4.5	84	0.00
32 T	1,1,1-Trichloroethane	0.969	1.018	-5.1	91	0.00
33 S	1,2-Dichloroethane-d4	0.719	0.613	14.7	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.445	0.406	8.8	80	0.00
36 T	1,1-Dichloropropene	0.636	0.613	3.6	85	0.00
37 T	Ethyl Acetate	0.697	0.679	2.6	86	0.00
38 T	Carbon Tetrachloride	0.678	0.711	-4.9	89	0.00
39 T	Methylcyclohexane	0.752	0.710	5.6	82	0.00
40 TM	Benzene	1.858	1.896	-2.0	89	0.00
41 T	Methacrylonitrile	0.399	0.403	-1.0	91	0.00
42 TM	1,2-Dichloroethane	0.714	0.721	-1.0	89	0.00
43 T	Isopropyl Acetate	1.070	1.093	-2.1	87	0.00
44 TM	Trichloroethene	0.568	0.574	-1.1	90	0.00
45 C	1,2-Dichloropropane	0.463	0.482	-4.1#	92	0.00

Data Path : W:\HPCHEM1\MSVOA X\DATA\WX051218\
 Data File : VX001678.D
 Acq On : 12 May 2018 20:16
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 14 07:24:19 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.328	0.331	-0.9	88	0.00
47 T	Bromodichloromethane	0.607	0.665	-9.6	92	0.00
48 T	Methyl methacrylate	0.583	0.591	-1.4	86	0.00
49 T	1,4-Dioxane	0.013	0.012	7.7	84	0.00
50 S	Toluene-d8	1.623	1.423	12.3	75	0.00
51 T	4-Methyl-2-Pentanone	0.711	0.742	-4.4	89	0.00
52 CM	Toluene	1.215	1.266	-4.2#	90	0.00
53 T	t-1,3-Dichloropropene	0.686	0.749	-9.2	88	0.00
54 T	cis-1,3-Dichloropropene	0.624	0.680	-9.0	87	0.00
55 T	1,1,2-Trichloroethane	0.489	0.505	-3.3	91	0.00
56 T	Ethyl methacrylate	0.705	0.726	-3.0	86	0.00
57 T	1,3-Dichloropropane	0.809	0.825	-2.0	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.346	0.370	-6.9	87	0.00
59 T	2-Hexanone	0.545	0.541	0.7	84	0.00
60 T	Dibromochloromethane	0.501	0.547	-9.2	87	0.00
61 T	1,2-Dibromoethane	0.514	0.526	-2.3	87	0.00
62 S	4-Bromofluorobenzene	0.623	0.530	14.9	72	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.697	0.700	-0.4	84	0.00
65 PM	Chlorobenzene	1.466	1.495	-2.0	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.525	0.563	-7.2	87	0.00
67 C	Ethyl Benzene	2.403	2.451	-2.0#	85	0.00
68 T	m/p-Xylenes	0.950	0.979	-3.1	85	0.00
69 T	o-Xylene	0.923	0.958	-3.8	85	0.00
70 T	Styrene	1.499	1.601	-6.8	86	0.00
71 P	Bromoform	0.430	0.470	-9.3	85	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.461	3.584	-3.6	85	0.00
74 T	N-amyl acetate	1.572	1.625	-3.4	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.022	1.074	-5.1	89	0.00
76 T	1,2,3-Trichloropropane	1.014	1.044	-3.0	87	0.00
77 T	Bromobenzene	1.006	1.014	-0.8	86	0.00
78 T	n-propylbenzene	3.769	3.975	-5.5	85	0.00
79 T	2-Chlorotoluene	2.320	2.388	-2.9	85	0.00
80 T	1,3,5-Trimethylbenzene	2.882	2.992	-3.8	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.266	0.281	-5.6	80	0.00
82 T	4-Chlorotoluene	2.676	2.753	-2.9	84	0.00
83 T	tert-Butylbenzene	2.867	2.981	-4.0	84	0.00
84 T	1,2,4-Trimethylbenzene	2.962	3.090	-4.3	84	0.00
85 T	sec-Butylbenzene	3.442	3.664	-6.4	86	0.00
86 T	p-Isopropyltoluene	3.157	3.327	-5.4	84	0.00
87 T	1,3-Dichlorobenzene	1.799	1.829	-1.7	85	0.00
88 T	1,4-Dichlorobenzene	1.835	1.832	0.2	84	0.00
89 T	n-Butylbenzene	2.616	2.683	-2.6	82	0.00

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001678.D
Acq On : 12 May 2018 20:16
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 14 07:24:19 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.475	0.539	-13.5	90	0.00
91 T	1,2-Dichlorobenzene	1.830	1.887	-3.1	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.225	0.240	-6.7	84	0.00
93 T	1,2,4-Trichlorobenzene	1.339	1.398	-4.4	84	0.00
94 T	Hexachlorobutadiene	0.733	0.734	-0.1	84	0.00
95 T	Naphthalene	3.704	4.027	-8.7	85	0.00
96 T	1,2,3-Trichlorobenzene	1.374	1.441	-4.9	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6