

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001671.D
 Acq On : 12 May 2018 17:09
 Operator : JC/MD
 Sample : J2741-07DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW301-180502DL

Quant Time: May 14 07:15:59 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264863	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	356360	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	310379	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	165507	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	211831	55.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.28%	
35) Dibromofluoromethane	5.51	113	169629	53.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.02%	
50) Toluene-d8	8.72	98	591865	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	11.14	95	191137	43.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.06%	
Target Compounds						
4) Vinyl Chloride	1.40	62	25620	7.07	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	254565	73.08	ug/l	86
64) Tetrachloroethene	9.34	164	1473	0.34	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001671.D
Acq On : 12 May 2018 17:09
Operator : JC/MD
Sample : J2741-07DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW301-180502DL

Quant Time: May 14 07:15:59 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.00 min

Lab File: VX001671.D

Acq: 12 May 2018 17:09

Instrument :

MSVOA_X

ClientSampleId :


SS037MW301-180502DL

Tot Ion:168 Resp: 264863

Ion Ratio Lower Upper




168 100

99 45.8 36.6 54.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#4

Vinyl Chloride

Concen: 7.07 ug/l

RT: 1.40 min Scan# 134

Delta R.T. 0.00 min

Lab File: VX001671.D


Acq: 12 May 2018 17:09

Tgt Ion: 62 Resp: 25620

Ion Ratio Lower Upper



62 100


64 31.0 25.8 38.6



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 63.90 (63.60 to 64.60): VX0





#27

cis-1,2-Dichloroethene

Concen: 73.08 ug/l

RT: 4.60 min Scan# 659

Delta R.T. 0.00 min

Lab File: VX001671.D

Acq: 12 May 2018 17:09

Instrument :

MSVOA_X

ClientSampleId :

SS037MW301-180502DL


Tot Ion: 96 Resp: 254565

Ion Ratio Lower Upper




96 100

61 126.3 0.0 301.6

98 64.5 0.0 131.6



Abundance Ion 95.90 (95.60 to 96.60): VX0
Ion 60.90 (60.60 to 61.60): VX0
Ion 97.90 (97.60 to 98.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 55.64 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.00 min

Lab File: VX001671.D


Acq: 12 May 2018 17:09

Tgt Ion: 65 Resp: 211831



Ion Ratio Lower Upper


65 100

67 50.5 0.0 101.8



Abundance Ion 64.90 (64.60 to 65.60): VX0
Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.87 min Scan# 1030
 Delta R.T. 0.00 min
 Lab File: VX001671.D
 Acq: 12 May 2018 17:09

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW301-180502DL

Tot Ion	Ratio	Lower	Upper
114	100		
63	18.9	0.0	38.6
88	14.6	0.0	30.8




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX
 Ion 87.90 (87.60 to 88.60): VX






#35
 Dibromofluoromethane
 Concen: 53.51 ug/l
 RT: 5.51 min Scan# 807
 Delta R.T. 0.00 min
 Lab File: VX001671.D
 Acq: 12 May 2018 17:09

Tgt Ion	Ratio	Lower	Upper
113	100		
111	102.2	82.2	123.2
192	26.6	21.4	32.0



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 51.15 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX001671.D

Acq: 12 May 2018 17:09

Instrument :

MSVOA_X

ClientSampleId :


SS037MW301-180502DL

Tot Ion: 98 Resp: 591865

Ion Ratio Lower Upper




98 100

100 66.4 54.0 81.0



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V



#62

4-Bromofluorobenzene

Concen: 43.03 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX001671.D

Acq: 12 May 2018 17:09


Tgt Ion: 95 Resp: 191137

Ion Ratio Lower Upper

95 100

174 102.1 0.0 202.0



176 98.8 0.0 197.4




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.01 min
Lab File: VX001671.D
Acq: 12 May 2018 17:09

Instrument :
MSVOA_X
ClientSampleId :
SS037MW301-180502DL

Tot Ion:117 Resp: 310379
Ion Ratio Lower Upper
117 100
82 53.7 40.0 60.0
119 32.6 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#64
Tetrachloroethene
Concen: 0.34 ug/l
RT: 9.34 min Scan# 1436
Delta R.T. 0.00 min
Lab File: VX001671.D
Acq: 12 May 2018 17:09

Tgt Ion:164 Resp: 1473
Ion Ratio Lower Upper
164 100
166 145.7 103.5 155.3
129 90.6 66.7 100.1
131 74.0 64.9 97.3



Abundance Ion 163.80 (163.50 to 164.50): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.08 min Scan# 1886

Delta R.T. 0.00 min

Lab File: VX001671.D

Acq: 12 May 2018 17:09

Instrument :

MSVOA_X

ClientSampleId :

SS037MW301-180502DL


Tot Ion:152 Resp: 165507

Ion Ratio Lower Upper

152 100

115 51.9 37.7 113.1

150 155.8 0.0 349.4



Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

