

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001681.D
 Acq On : 12 May 2018 22:02
 Operator : JC/MD
 Sample : VX0512WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0512WBL02

Quant Time: May 14 07:26:18 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	273037	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	369730	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	315584	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	165463	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	175655	44.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.52%	
35) Dibromofluoromethane	5.51	113	142514	43.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.66%	
50) Toluene-d8	8.72	98	504264	42.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.02%	
62) 4-Bromofluorobenzene	11.14	95	155756	33.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	67.60%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001681.D
Acq On : 12 May 2018 22:02
Operator : JC/MD
Sample : VX0512WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0512WBL02

Quant Time: May 14 07:26:18 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.00 min

Lab File: VX001681.D

Acq: 12 May 2018 22:02

Instrument :

MSVOA_X

ClientSampleId :


VX0512WBL02

Tot Ion:168 Resp: 273037

Ion Ratio Lower Upper




168 100

99 47.3 36.6 54.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 44.76 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.00 min

Lab File: VX001681.D


Acq: 12 May 2018 22:02

Tgt Ion: 65 Resp: 175655

Ion Ratio Lower Upper



65 100


67 51.3 0.0 101.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VX001681.D

Acq: 12 May 2018 22:02

Instrument :

MSVOA_X

ClientSampleId :

VX0512WBL02


Tot Ion:114 Resp: 369730

Ion Ratio Lower Upper

114 100

63 18.3 0.0 38.6




88 14.7 0.0 30.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 43.33 ug/l

RT: 5.51 min Scan# 807

Delta R.T. 0.00 min

Lab File: VX001681.D

Acq: 12 May 2018 22:02


Tgt Ion:113 Resp: 142514

Ion Ratio Lower Upper

113 100

111 101.8 82.2 123.2



192 26.5 21.4 32.0




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 42.01 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX001681.D

Acq: 12 May 2018 22:02

Instrument :

MSVOA_X

ClientSampleId :


VX0512WBL02

Tot Ion: 98 Resp: 504264

Ion Ratio Lower Upper




98 100

100 66.3 54.0 81.0



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 33.80 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX001681.D

Acq: 12 May 2018 22:02


Tgt Ion: 95 Resp: 155756

Ion Ratio Lower Upper

95 100

174 101.6 0.0 202.0



176 99.2 0.0 197.4




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.01 min
Lab File: VX001681.D
Acq: 12 May 2018 22:02

Instrument :
MSVOA_X
ClientSampleId :
VX0512WBL02

Tot Ion:117 Resp: 315584
Ion Ratio Lower Upper
117 100
82 51.8 40.0 60.0
119 32.2 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. 0.00 min
Lab File: VX001681.D
Acq: 12 May 2018 22:02

Tgt Ion:152 Resp: 165463
Ion Ratio Lower Upper
152 100
115 52.3 37.7 113.1
150 155.9 0.0 349.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

