

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
 Data File : VX001685.D
 Acq On : 12 May 2018 23:48
 Operator : JC/MD
 Sample : PB109200ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB109200ZHE#01

Quant Time: May 14 07:31:30 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
 Quant Title : SW846 8260
 QLast Update : Sun May 06 07:59:25 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	267440	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	370029	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	326039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	177609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	178314	46.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.78%	
35) Dibromofluoromethane	5.51	113	142885	43.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.82%	
50) Toluene-d8	8.72	98	513910	42.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.56%	
62) 4-Bromofluorobenzene	11.14	95	164114	35.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.16%	
Target Compounds						
16) Acetone	2.45	43	6631	3.80	ug/l	Qvalue 100
18) Methyl Acetate	2.78	43	21709	4.73	ug/l	100
20) Methylene Chloride	2.86	84	2525	0.73	ug/l #	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051218\
Data File : VX001685.D
Acq On : 12 May 2018 23:48
Operator : JC/MD
Sample : PB109200ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB109200ZHE#01

Quant Time: May 14 07:31:30 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X050518W.M
Quant Title : SW846 8260
QLast Update : Sun May 06 07:59:25 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.00 min

Lab File: VX001685.D

Acq: 12 May 2018 23:48

Instrument :

MSVOA_X

ClientSampleId :


PB109200ZHE#01

Tot Ion:168 Resp: 267440

Ion Ratio Lower Upper




168 100

99 48.0 36.6 54.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#16

Acetone

Concen: 3.80 ug/l

RT: 2.45 min Scan# 305

Delta R.T. -0.01 min

Lab File: VX001685.D


Acq: 12 May 2018 23:48

Tgt Ion: 43 Resp: 6631

Ion Ratio Lower Upper



43 100


58 29.6 23.8 35.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#18
Methvl Acetate
Concen: 4.73 ug/l
RT: 2.78 min Scan# 360
Delta R.T. 0.00 min
Lab File: VX001685.D
Acq: 12 May 2018 23:48




Instrument :
MSVOA_X
ClientSampleId :
PB109200ZHE#01


Tot Ion: 43 Resp: 21709
Ion Ratio Lower Upper
43 100
74 23.2 18.6 27.8



#20
Methylene Chloride
Concen: 0.73 ug/l
RT: 2.86 min Scan# 372
Delta R.T. -0.01 min
Lab File: VX001685.D
Acq: 12 May 2018 23:48

Tgt Ion: 84 Resp: 2525
Ion Ratio Lower Upper
84 100
49 114.5 96.6 144.8
51 47.4 29.0 43.4#
86 62.7 52.2 78.4









#33
 1,2-Dichloroethane-d4
 Concen: 46.39 ug/l
 RT: 6.07 min Scan# 900
 Delta R.T. 0.00 min
 Lab File: VX001685.D
 Acq: 12 May 2018 23:48




Instrument :
 MSVOA_X
 ClientSampleId :
 PB109200ZHE#01


Tot Ion: 65 Resp: 178314
 Ion Ratio Lower Upper
 65 100
 67 50.4 0.0 101.8



#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.87 min Scan# 1030
 Delta R.T. 0.00 min
 Lab File: VX001685.D
 Acq: 12 May 2018 23:48

Tgt Ion: 114 Resp: 370029
 Ion Ratio Lower Upper
 114 100
 63 19.2 0.0 38.6
 88 15.0 0.0 30.8









#35
 Dibromofluoromethane
 Concen: 43.41 ug/l
 RT: 5.51 min Scan# 807
 Delta R.T. 0.00 min
 Lab File: VX001685.D
 Acq: 12 May 2018 23:48




Instrument :
 MSVOA_X
 ClientSampleId :
 PB109200ZHE#01


Tot Ion:113 Resp: 142885
 Ion Ratio Lower Upper
 113 100
 111 103.0 82.2 123.2
 192 26.9 21.4 32.0



#50
 Toluene-d8
 Concen: 42.78 ug/l
 RT: 8.72 min Scan# 1334
 Delta R.T. 0.00 min
 Lab File: VX001685.D
 Acq: 12 May 2018 23:48

Tgt Ion: 98 Resp: 513910
 Ion Ratio Lower Upper
 98 100
 100 66.1 54.0 81.0









#62
4-Bromofluorobenzene
Concen: 35.58 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. 0.00 min
Lab File: VX001685.D
Acq: 12 May 2018 23:48




Instrument :
MSVOA_X
ClientSampleId :
PB109200ZHE#01


Tot Ion: 95 Resp: 164114
Ion Ratio Lower Upper
95 100
174 100.9 0.0 202.0
176 99.1 0.0 197.4



#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.01 min
Lab File: VX001685.D
Acq: 12 May 2018 23:48

Tgt Ion: 117 Resp: 326039
Ion Ratio Lower Upper
117 100
82 52.4 40.0 60.0
119 33.0 25.8 38.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.09 min Scan# 1886

Delta R.T. 0.00 min

Lab File: VX001685.D

Acq: 12 May 2018 23:48

Instrument :

MSVOA_X

ClientSampleId :

PB109200ZHE#01


Tot Ion:152 Resp: 177609

Ion Ratio Lower Upper

152 100

115 52.0 37.7 113.1

150 155.9 0.0 349.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

