

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
 Data File : VX001724.D
 Acq On : 14 May 2018 22:16
 Operator : JC/MD
 Sample : J2833-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: May 16 10:41:22 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	375893	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	332228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	198887	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	173311	46.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.30%	
35) Dibromofluoromethane	5.51	113	141243	44.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.06%	
50) Toluene-d8	8.72	98	517966	45.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.18%	
62) 4-Bromofluorobenzene	11.14	95	172429	42.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.30%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
Data File : VX001724.D
Acq On : 14 May 2018 22:16
Operator : JC/MD
Sample : J2833-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: May 16 10:41:22 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

