

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX051518\
 Data File : VX001739.D
 Acq On : 15 May 2018 12:27
 Operator : JC/MD
 Sample : J2737-08 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW128-180503

Quant Time: May 16 04:44:00 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	269553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	362419	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	299736	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	155556	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	194494	53.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.66%	
35) Dibromofluoromethane	5.51	113	160266	52.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.80%	
50) Toluene-d8	8.72	98	570772	52.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.22%	
62) 4-Bromofluorobenzene	11.14	95	171634	43.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.04%	
Target Compounds						
4) Vinyl Chloride	1.40	62	131078	41.493	ug/l	99
21) trans-1,2-Dichloroethene	3.17	96	6870	2.341	ug/l	92
27) cis-1,2-Dichloroethene	4.60	96	1163833	348.477	ug/l	90
44) Trichloroethene	7.22	130	645297	169.137	ug/l	100
64) Tetrachloroethene	9.34	164	162964	41.765	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_X\DATA\VX051518\
Data File : VX001739.D
Acq On : 15 May 2018 12:27
Operator : JC/MD
Sample : J2737-08 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW128-180503

Quant Time: May 16 04:44:00 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

