

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
 Data File : VX001741.D
 Acq On : 15 May 2018 13:21
 Operator : JC/MD
 Sample : J2737-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3TB06-180503

Quant Time: May 16 11:08:17 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	267719	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	358446	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	292335	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	148994	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	190020	52.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.90%	
35) Dibromofluoromethane	5.51	113	156168	51.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.26%	
50) Toluene-d8	8.72	98	551437	50.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.80%	
62) 4-Bromofluorobenzene	11.15	95	164204	42.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
Data File : VX001741.D
Acq On : 15 May 2018 13:21
Operator : JC/MD
Sample : J2737-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3TB06-180503

Quant Time: May 16 11:08:17 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

