

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
 Data File : VX001743.D
 Acq On : 15 May 2018 14:14
 Operator : JC/MD
 Sample : J2737-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW139-180503

Quant Time: May 16 04:56:39 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	264034	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	347940	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	297450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	156616	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	189549	53.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.12%	
35) Dibromofluoromethane	5.51	113	156601	53.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.66%	
50) Toluene-d8	8.72	98	531055	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
62) 4-Bromofluorobenzene	11.14	95	171741	45.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.72%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1137	0.37	ug/l	91
21) trans-1,2-Dichloroethene	3.17	96	1676	0.58	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	62199	19.01	ug/l	93
44) Trichloroethene	7.22	130	36825	10.05	ug/l	98
64) Tetrachloroethene	9.34	164	543463	140.35	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
Data File : VX001743.D
Acq On : 15 May 2018 14:14
Operator : JC/MD
Sample : J2737-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW139-180503

Quant Time: May 16 04:56:39 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

