

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
 Data File : VX001747.D
 Acq On : 15 May 2018 16:02
 Operator : JC/MD
 Sample : J2737-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0454-180503

Quant Time: May 16 05:22:05 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	243565	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	325298	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	276607	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	142869	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	176141	53.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.90%	
35) Dibromofluoromethane	5.51	113	144017	52.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.92%	
50) Toluene-d8	8.72	98	479518	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
62) 4-Bromofluorobenzene	11.15	95	153373	43.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.66%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	6652	2.20	ug/l	84
44) Trichloroethene	7.22	130	3610	1.05	ug/l	95
64) Tetrachloroethene	9.34	164	24454	6.79	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
Data File : VX001747.D
Acq On : 15 May 2018 16:02
Operator : JC/MD
Sample : J2737-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0454-180503

Quant Time: May 16 05:22:05 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

