

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
 Data File : VX001748.D
 Acq On : 15 May 2018 16:29
 Operator : JC/MD
 Sample : J2856-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-DUPLICATE

Quant Time: May 16 11:26:23 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251839	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	337722	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	288438	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	149806	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161844	47.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.90%	
35) Dibromofluoromethane	5.51	113	132405	46.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.92%	
50) Toluene-d8	8.72	98	441643	43.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.54%	
62) 4-Bromofluorobenzene	11.15	95	141047	38.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.76%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
Data File : VX001748.D
Acq On : 15 May 2018 16:29
Operator : JC/MD
Sample : J2856-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-DUPLICATE

Quant Time: May 16 11:26:23 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

