

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
 Data File : VX001754.D
 Acq On : 15 May 2018 19:10
 Operator : JC/MD
 Sample : J2796-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS033-180507

Quant Time: May 16 05:42:41 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	231052	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	311929	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	262458	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131933	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	178615	57.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.34%	
35) Dibromofluoromethane	5.51	113	146041	55.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.96%	
50) Toluene-d8	8.72	98	480543	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
62) 4-Bromofluorobenzene	11.15	95	149461	44.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	21005	17.27	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	1930	0.67	ug/l	# 18
64) Tetrachloroethene	9.34	164	1786	0.52	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
Data File : VX001754.D
Acq On : 15 May 2018 19:10
Operator : JC/MD
Sample : J2796-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS033-180507

Quant Time: May 16 05:42:41 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

