

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
 Data File : VX001757.D
 Acq On : 15 May 2018 20:31
 Operator : JC/MD
 Sample : J2737-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW146-180503

Quant Time: May 16 04:30:10 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	237256	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	316823	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	271631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	139289	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	173032	54.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.82%	
35) Dibromofluoromethane	5.51	113	144010	53.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.72%	
50) Toluene-d8	8.72	98	486688	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	11.15	95	153046	44.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.78%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	35859	12.90	ug/l	100
12) 1,1-Dichloroethene	2.38	96	2595	1.11	ug/l	96
21) trans-1,2-Dichloroethene	3.17	96	4625	1.79	ug/l	91
27) cis-1,2-Dichloroethene	4.60	96	988721	336.34	ug/l	90
40) Benzene	6.14	78	2558	0.24	ug/l	99
44) Trichloroethene	7.22	130	336434	100.87	ug/l	99
64) Tetrachloroethene	9.35	164	4199894	1187.73	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
Data File : VX001757.D
Acq On : 15 May 2018 20:31
Operator : JC/MD
Sample : J2737-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW146-180503

Quant Time: May 16 04:30:10 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

