


Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051518\
 Data File : VX001734.D
 Acq On : 15 May 2018 09:34
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Title : SW846 8260
 Last Update : Tue May 15 01:36:22 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1723

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	4097	PASS
75	95	30	60	47.2	10904	PASS
95	95	100	100	100.0	23122	PASS
96	95	5	9	6.4	1483	PASS
173	174	0.00	2	1.0	230	PASS
174	95	50	100	98.4	22755	PASS
175	174	5	9	7.6	1720	PASS
176	174	95	101	95.4	21697	PASS
177	176	5	9	6.5	1412	PASS