

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051618\
 Data File : VX001774.D
 Acq On : 16 May 2018 19:37
 Operator : JC/MD
 Sample : J2853-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW214-180507

Quant Time: May 17 04:21:59 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 15 01:36:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	268737	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	361983	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	289407	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	139580	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	213407	59.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.48%	
35) Dibromofluoromethane	5.51	113	177555	58.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.24%	
50) Toluene-d8	8.72	98	623401	56.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.96%	
62) 4-Bromofluorobenzene	11.15	95	173179	43.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.92%	
Target Compounds						
16) Acetone	2.45	43	2673	1.89	ug/l	93
44) Trichloroethene	7.22	130	1698	0.45	ug/l	100
64) Tetrachloroethene	9.34	164	10237	2.72	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX051618\
Data File : VX001774.D
Acq On : 16 May 2018 19:37
Operator : JC/MD
Sample : J2853-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW214-180507

Quant Time: May 17 04:21:59 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X051518W.M
Quant Title : SW846 8260
QLast Update : Tue May 15 01:36:22 2018
Response via : Initial Calibration

